



# E. I. du Pont de Nemours and Company

Wilmington, Delaware 19880-0024

## Certificate of Sterility - Gamma Irradiation

<b>Sterility Run Number:</b>	<b>M1600946B</b>
<b>Product Release Date:</b>	<b>05/03/2016 22:05:44</b>
<b>Minimum Specified Dose (kGy):</b>	<b>25.0</b>
<b>Maximum Specified Dose (kGy):</b>	<b>35.0</b>
<b>Minimum Validation Dose (kGy):</b>	<b>30.0</b>
<b>Minimum Delivered Dose (kGy):</b>	<b>34.9</b>
<b>Maximum Delivered Dose (kGy):</b>	<b>51.6</b>

<b>Product Number:</b>	<b>IC183BWHMD0025DS</b>
<b>Product Name:</b>	<b>DUPONT. TYVEK. ISOCLEAN. COVERALL</b>
<b>Product Size:</b>	<b>MD</b>
<b>Manufacturing Lot Number:</b>	<b>AHS1614602</b>
<b>Sterilization Lot Number:</b>	<b>AHS1614602</b>

*DuPont certifies that the product number and lot numbers above have been processed and released according to DuPont specifications and standard operating procedures. The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products - Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.*

*Certificates of Sterility can be retrieved on-line at [www.personalprotection.dupont.com](http://www.personalprotection.dupont.com) or contact DuPont Customer Service at 1-800-931-3456.*

*Approved By: Edward K. Mitchell  
Quality Assurance*

E. I. du Pont de Nemours and Company

This report is provided at no charge solely to assist DuPont customers in meeting their quality systems needs. The recipient of this report agrees to evaluate such information, to make an independent decision regarding the suitability of such information for its application, and only use such information pursuant to their current good product stewardship principles and all regulatory requirements applicable to your business.